**NOVEL CYLINDRICAL/SPHERICAL SCANNER CONFIGURATION FOR
REVERSE ENGINEERING**

by

**Medhat Samaan
B Sc. Electrical Engineering, Ain Shams University,
Cairo, Egypt, 1980**

**A thesis Presented to the Department of Mechanical Engineering
in partial fulfillment of the requirements
for the degree of Master of Applied Science**

DEC 13 2004

**Ryerson University
Toronto, Ontario, Canada, 2004**

**PROPERTY OF
Ryerson University Library**

Copyright © Medhat Samaan, 2004

UMI Number: EC52975

All rights reserved

INFORMATION TO USERS

The quality of this reproduction is dependent upon the quality of the copy submitted. Broken or indistinct print, colored or poor quality illustrations and photographs, print bleed-through, substandard margins, and improper alignment can adversely affect reproduction.

In the unlikely event that the author did not send a complete manuscript and there are missing pages, these will be noted. Also, if unauthorized copyright material had to be removed, a note will indicate the deletion.

UMI[®]

UMI Microform EC52975

Copyright 2008 by ProQuest LLC

All rights reserved. This microform edition is protected against unauthorized copying under Title 17, United States Code.

ProQuest LLC
789 East Eisenhower Parkway
P.O. Box 1346
Ann Arbor, MI 48106-1346

Author's Declaration

I hereby declare that I am the sole author of this thesis.

I authorize Ryerson University to lend this thesis to other institutions or individuals for the purpose of scholarly research.

~~Medhat~~ Samaan

I further authorize Ryerson University to reproduce this thesis by photocopying or by other means, in total or in part, at the request of other institutions or individuals for the purpose of scholarly research.

~~Medhat~~ Samaan

Ryerson University requires the signatures of all persons using or photocopying this thesis. Please sign below, and give address and date.

[illegible]

Abstract

Novel Cylindrical/Spherical Scanner Configuration for Reverse Engineering

© Medhat Samaan, 2004

Master of Applied Science

Department of Mechanical and Industrial Engineering

Ryerson University

Laser scanners are often used in geometric reverse engineering to generate a CAD drawing out of an existing physical part. This often requires careful path planning to ensure the correct stand-off distance of the sensor, to prevent any collisions of the sensor with the target object and to prevent any scan occlusions in the resulting data.

A dedicated system was constructed to collect data using a 3-axis cylindrical/spherical device. The benefit of this scanning system is the ability to scan completely around the target object, including the bottom of the object. A stripe-type laser diode is used along with a CCD camera. The target object sits

on a transparent plexiglass table, which can be rotated 360° . Both the laser and the CCD are mounted on an arm, such that they can be traversed along a linear path, as well as being rotated 360° around the table. This permits the device to scan in either spherical or cylindrical paths with 360° rotation around the target object. This system allows for the quick scanning of any object while minimizing the number of occlusions in the resulting scan data.

PROPERTY OF
Ryerson University Library

Acknowledgements

I am especially grateful to my supervisor Dr. Vincent H. Chan of the Department of Mechanical Engineering for his knowledgeable advice, support, and encouragement during the course of this research. This work would not have been possible without his help.

I'd also like to thank Deven Ostrom, Technical Officer, of the Department of Mechanical Engineer for his assistance.

Last but not least I'd like to express my appreciation to all Ryerson faculty and staff for the knowledge I gained as well as the friendly atmosphere I enjoyed during the period of my study.

Table of Contents

Author's Declaration	ii
Borrower's Page	iii
Abstract	iv
Acknowledgements	vi
Table of Contents	vii
List of Figures	viii
List of Tables	x
List of Equations	x
 Chapter 1 Introduction	 1
1.1 Literature Review	4
1.2 Proposed New Scanner	21
 Chapter 2 Construction of the Proposed Scanner	 23
2.1 Control of the Scanner	26
2.2 Scanning Sensor	28
2.3 Conversion of Pixel Values to XYZ Coordinates	32
2.4 Transformation for the Rotary Table and Arm	35
2.5 Call Interface Node (CIN)	37
2.6 DXF Format	40
2.7 Master Algorithm	42
 Chapter 3 Preliminary Result of the Prototype	 45
3.1 Laser Alignment and Separation	46
3.2 Motor Control	47
3.3 CIN Node	48
3.4 Image Processing	49
3.5 Full Scan Results	54
3.5.1 Cylindrical Scan	54
3.5.2 Spherical Scan	63
3.6 Calibration	65
3.7 Comparison with Other Machines	66
 Chapter 4 Conclusions and Future Work	 67
 References	 69
 Appendix A Build the CIN Node	 71
Appendix B C Program for the CIN Node (Cylindrical)	72
Appendix C Full LabVIEW Program for Cylindrical Scan	75
Appendix D C Program for the CIN Node (Spherical)	84
Appendix E Full LabVIEW Program for Spherical Scan	86
Appendix F Technology Comparison Chart	94

List of Figures

Fig 1.1	CCD directed scanning [1] -----	6
Fig 1.2	Laser light stripe projection [3] -----	7
Fig 1.3	Adjustable light displacement system [4] -----	8
Fig 1.4	Raster Scanning device [5] -----	9
Fig 1.5	Mirror synchronized scanner [6] -----	10
Fig.1.6	Orthogonal laser scanner [8] -----	11
Fig 1.7	CMM mounted structured light measuring system [9] -----	12
Fig 1.8	Detail of camera orientation [9] -----	12
Fig 1.9	Fixed camera/rotary table machine vision system [13]-----	14
Fig 1.10	Stereo laser line reconstruction [14] -----	15
Fig 1.11	Linear and rotary laser scanner [15] -----	16
Fig 1.12	Top view, turn-table laser scanner [16] -----	17
Fig 1.13	Side view, turn-table laser scanner [16] -----	17
Fig 1.14	Rotary scanner, showing voids [17] -----	18
Fig 1.15	Rotary scanner showing (a) first view (b) overlap [17] -----	18
Fig 1.16	Three Cartesian axis scanner [18] -----	19
Fig 1.17	Rotary scanner [18] -----	20
Fig 1.18	Robotic arm scanner [18] -----	20
Fig 1.19	Proposed scanner for orthogonal, cylindrical and spherical motions -----	22
Fig 2.1	AutoCAD drawing of the prototype -----	24
Fig 2.2	A photo of the constructed scanner -----	25
Fig 2.3	PC interfacing with the scanner components -----	26
Fig 2.4	Motor control box -----	26
Fig 2.5	Motor drive LabVIEW program (front panel) -----	27
Fig 2.6	Motor drive LabVIEW program (block diagram) -----	27
Fig 2.7	Image capture LabVIEW program (front panel) -----	29
Fig 2.8	Image capture LabVIEW program (block diagram) -----	29
Fig 2.9	Focus adjustment using graph paper -----	30
Fig 2.10	Adjustment of laser line illumination and alignment -----	31
Fig 2.13	Camera/laser sled -----	32
Fig 2.14	Laser camera and object triangle -----	33
Fig 2.15	(.c) file generated by CIN node for cylindrical scan -----	38
Fig 2.16	(.c) file generated by CIN node for spherical scan -----	39
Fig 2.17	Beginning of DXF file format -----	40

Fig 2.18	C code to generate the beginning of DXF file -----	41
Fig 2.19	Format for displaying a point by DXF file -----	41
Fig 2.20	C code to generate a point by the DXF file -----	41
Fig 2.21	End of DXF file format -----	41
Fig 2.22	C code to generate the end of DXF file -----	41
Fig 2.23	Cylindrical scan algorithm flowchart -----	43
Fig 2.24	Spherical scan algorithm flowchart -----	44
Fig 3.1	Object with quadric and free form surfaces -----	46
Fig 3.2	Reflection caused by shiny gage block -----	49
Fig 3.3	Gage scan with background reflection from the rotary table -	50
Fig 3.4	Gage scan and reflection caused by the rotary table -----	51
Fig 3.5	Object scan and reflection caused by the rotary table -----	51
Fig 3.6	1-2-3-gage block scan with room light reflection -----	51
Fig 3.7	Cast object scan with room light reflection -----	52
Fig 3.8	Bottom scan with very low illumination -----	52
Fig 3.9	Bottom scan with high illumination -----	53
Fig 3.10	Skew scan due to unsupported rotary table -----	53
Fig 3.11	Side scan, before applying the transformation matrix -----	53
Fig 3.12	The machine adjusted for the top scan -----	54
Fig 3.13	Top scan of a 1-2-3-gage block, where $Z = 1$ -----	55
Fig 3.14	Top scan of a 1-2-3-gage block, where $Z = 3$ -----	55
Fig 3.15	Top scan while the rotary table moves 90° -----	55
Fig 3.16	Top scan while the rotary table moves 180° -----	56
Fig 3.17	Top scan while the rotary table moves 270° -----	56
Fig 3.18	Top scan for cast object with quadric and free form surfaces	56
Fig 3.19	The machine adjusted for the side scan -----	57
Fig 3.20	Side scan for 1-2-3-gage block while the rotary table at 0° --	58
Fig 3.21	Side scan for 1-2-3-gage block while the rotary table at 90°	58
Fig 3.22	Side scan for 1-2-3-gage block while the rotary table at 180°	59
Fig 3.23	Side scan for 1-2-3-gage block while the rotary table at 270°	59
Fig 3.24	Side scan for cast object while the rotary table at 0° -----	59
Fig 3.25	Side scan for cast object while the rotary table at 90° -----	60
Fig 3.26	Side scan for cast object while the rotary table at 180° -----	60
Fig 3.27	Side scan for cast object while the rotary table at 270° -----	60
Fig 3.28	The machine adjusted for the bottom scan -----	61
Fig 3.29	Bottom scan for 1-2-3-gage block -----	62
Fig 3.30	Bottom scan for cast object -----	62
Fig 3.31	Spherical scan for 1-2-3-gage block while the arm at 30° ----	63
Fig 3.32	Spherical scan for 1-2-3-gage block while the arm at 60° ----	63
Fig 3.33	Spherical scan for 1-2-3-gage block while the arm at 90° ----	64
Fig 3.34	Spherical scan for cast object while the arm at 60° -----	64
Fig 3.35	Spherical scan for cast object while the arm at 90° -----	64

List of Tables

Table 2.1 Mechanical Components of Scanner	24
--	----

List of Equations

Equation 1	Center of gravity calculation	31
Equation 2	Conversion of pixel value to Y coordinate	32
Equation 3	X / Z relation	33
Equation 4	Calculation for Z coordinate	33
Equation 5	Calculation for offset in X coordinate	33
Equation 6	Resolution in Z direction	33
Equation 7	Resolution in X offset	33
Equation 8	X-coordinate	34
Equation 9	Y-coordinate	34
Equation 10	Z-coordinate	34
Equation 11	Rotation around Z axis transformation matrix	35
Equation 12	Rotation around X axis transformation matrix	35
Equation 13	Total transformation matrix	35
Equation 14	X coordinate with respect to the machine reference frame ---	36
Equation 15	Y coordinate with respect to the machine reference frame ---	36
Equation 16	Z coordinate with respect to the machine reference frame ---	36
Equation 17	Rotation angle around X axis	36
Equation 18	Rotation angle around Z axis	36
Equation 19	Number of rows in an array format for CIN node	37
Equation 20	Number of columns in an array format for CIN node	37
Equation 21	Value of element in an array format for CIN node	37

Chapter 1

Introduction

Reverse engineering is the process of generating CAD models out of an existing physical part. The need for CAD models, taken from existing complex sculptured surfaces, is initiated because:

- 1- The design of a product starts with a clay model or other hand carved model.
- 2- CAD models are not available for many parts that were manufactured in the past.
- 3- The part has been modified from the original design to fit its position or to change its look.

To modify or reproduce such parts in a computer integrated manufacturing (CIM) environment, CAD models should be created. CAD models are created based on measured data from the part's surfaces. Reverse Engineering is used in industrial environments such as the manufacturing of car bodies, ship hulls, airplane fuselages and wings, propellers, shoe insoles, bottles, and so forth. The CAD models for these parts are often created based on digitized data captured from a physical prototype.

A fully automated reverse engineering process is still under research. Currently, the co-ordinate measuring machine (CMM) is manually used to guide a digitizing sensor to collect surface data points. This slow process, which might take hours to days, requires expensive equipment and well-trained professionals.

Due to the growing number of legacy parts, for which no CAD models exist, there is a growing interest in the representation of the surface geometry and topology of physical objects. Even the most trivial object requires multiple range images to be taken from different vantage points and integrated to form a complete model.

There are two main challenges to develop a fully automated reverse engineering system: the digitization of the scanned object image, and the conversion of the 3D scanned data into a compact form, compatible with CAD/CAM packages.

This thesis will deal with the digitization of a physical object. Machine configurations and conversion of 2D data into 3D are the two areas that have been developed in this work. To this end, a novel laser scanner configuration for reverse engineering has been developed. The scanner is able to:

1. Automate the reverse engineering system in terms of path planning
2. Scan with access from multiple viewpoints
3. Reduce scan occlusions
4. Scan the bottom of the part

1.1 Literature Review

Scanning plays an important role in reverse engineering since complex specifications have to be met for a greater number of products. Contact and non-contact methods are the two categories for scanning operations. The hardware used in sampling the point data from a part surface determines which category of scanning operations is used.

The most popular contact device is the CMM. These are numerical control (NC) driven and can be efficiently programmed to collect point samples for predefined features. CMMs have also brought a significant increase of productivity in performing scanning operations. They are accurate and efficient for checking dimensions and tolerances of a well-defined part but show difficulty in measuring parts with free-form surfaces. CMMs are inherently slow in acquiring point data since they need to make physical contact with a part surface for every point that is sampled. Moreover, CMMs can deform the part surface if the part is made of soft material. In industry, there is a great need to have high quality scan of parts with complicated and free-form surfaces. It is difficult and very time consuming to satisfy this need using traditional contact measuring devices.

The number of products with free-form surfaces is increasing. These are usually expensive parts and any occurrence of defects on the surface can cause costly errors in the final products or in the performance of the entire machine system. In order to perform a complete scan on parts with free-form surfaces, a scanning device needs to be able to quickly obtain a large amount of point data from the part surface.

In recent years, the accuracy of non-contact-type methods, compared to CMM, has been greatly improved and can provide a viable solution for this purpose. In non-contact devices the accessibility of scanned data is controlled by the boundaries of related surfaces with respect to the probe orientation.

Researchers have been working on machine configuration as it plays an important role in the digitization of the scanned object image. Contemporary research is moving towards the direction of line laser scanning using two methods of data collection. The first method employs a single camera while the second, known as stereovision, uses two cameras.

Chan, Bradley and Vickers [1] used a procedure where two scans were employed using a charge-coupled device (CCD) camera and CMM touch probe as shown in Fig 1.1. A CCD camera determines the location of the part on a CMM bed and recognizes the object's surface. The CCD scan is then used to determine the path for the second scan using a CMM touch probe or a laser scanner. The output of the second scan is then converted to CAD/CAM data using artificial neural networks.

Fig 1.1 CCD directed scanning [1]

In another work, Takeuchi, Shimizu and Mukai [2] set a TV camera over a CMM to take an image of the work area. A PC performs additional processing such as matching image data and path generation. The CMM is then directed to begin the measuring operation. Measured data from the touch probe of the CMM are sent back to the computer to evaluate the measurements.

Ozeki, Nakano and Yamamoto [3] used a device consisting of a TV camera, a light stripe projection unit, a light stripe scanning control unit, a signal-processing unit and a microcomputer as shown in Fig 1.2. This experimental set-up detects the range of an object by triangulation. The light stripe is projected onto the object while the TV camera receives the reflected image. The light stripe scanning control unit traverses the scanner over the whole top surface of the object.

Fig 1.2 Laser light stripe projection [3]

Jarvis [4] used a laser beam producing a sheet of light as shown in Fig 1.3. This light is scanned across the scene producing a single light stripe for each position. When the light source is displaced, the camera view of the stripe shows displacements, along a stripe, which is proportional to the depth. The proportionality constant between the beam displacement and depth is dependent upon the distance of the source from the camera. With this system more accurate depth measurements can be made with larger displacements.

Fig1.3 Adjustable light displacement system [4]

A block diagram of the experimental system, used by Harrison and Weir [5], is shown in Fig 1.4. An image acquisition board having a frame grabber of 256 x 256 pixels with 7-bit amplitude resolution was installed to capture images and to provide image output to a monitor. An attenuator wheel having seven attenuation steps was used to accommodate the very large reflectance range of the objects measured. Two translation tables provided steps in multiples of 0.5 mils, with a travel of 254 mm [10 inches] on the lower (y) table and 101 mm [4 inches] on the upper (x) table. With this system, all scanning was implemented by translating objects in a straight line or raster path, depending upon the mode of operation.

The angle between illuminating and receiving paths was set at about 30°, and the relative positions and angles of the receiving lens and camera were set experimentally to achieve the desired resolution of 2 mils per pixel at the object base plane, while maintaining a focal plane.

Fig1.4 Raster scanning device [5]

As shown in Fig 1.5, Feng, Liu and Xi [6], [7] used a double-sided scanning mirror to provide synchronization between projection and detection. The scanning mirror directs a laser beam onto the scanned surface. The diffuse laser light from the scanned surface reflects off the backside of the scanning mirror and is focused onto a CCD array. As the scanning mirror rotates, the incident laser beam sweeps out the scanning plane that intersects with the scanned object surface. From the position of the focused laser image point P on the CCD array, the 3D coordinate on the object's surface can be determined.

Fig 1.5 Mirror synchronized scanner [6]

In Maver and Bajcsy's [8] paper, the depth of the scene is measured by a camera-laser triangulation system, where the laser unit and the camera look toward the target with an acute angle as shown in Fig 1.6. A CCD camera and a laser unit are coupled and form a fixed sensor unit. The laser beam is spread into an illuminating plane. The illuminating plane intersects with the object surface, forming a laser stripe. A complete range image is obtained by scanning the scene with a series of parallel illuminating planes. This can be achieved by moving the scene support with a constant velocity perpendicular to the illuminating plane.

Fig 1.6 Orthogonal laser scanner [8]

Aguilar, Torres and Lope [9] have used two cameras attached to a CMM machine as shown in Fig 1.7. The components of the stereovision system are shown in Fig 1.8. The scan is made by projecting vertical light lines onto the object from a projector with a grating. The cameras detect the lines that cover the surface of the object. From the data collected by the cameras, a 3D model is established.

Fig 1.7 CMM mounted structured light measuring system [9]

Fig1.8 Detail of camera orientation [9]

Cho, Kim and Kim [10] used a vision camera mounted on the end effector of a CMM. Two images, which are needed for the 3D shape construction, are taken at two positions by moving the ram along the x-direction while the y-axis and z-coordinates of the camera remain unchanged. Using their surface recognition/localization method, the 3D surface is reconstructed.

Son, Kim and Lee [11] employed a laser scanner consisting of two major parts, the laser probe and the positioning system. The laser probe is mounted on a 3-axis transportation mechanism and moves along predetermined scan paths.

During scanning, a series of laser stripes is projected onto a surface patch and two CCD cameras detect the reflected beam. When the laser scanner captures an image, the system automatically finds an optical focus. The length of the laser stripe and the stand-off distance cannot be changed by an operator. Since the laser scanner consists of optical sensors and mechanical moving parts, various constraints must be satisfied when measuring a point on the part.

The 3D laser scanning mechanism employed in Lee, Park and Son's research [12] uses a laser stripe scanner that measures one laser stripe at a time. The main part of the laser scanning system is a laser probe, which is fixed on a 3-axis transport mechanism. The laser probe consists of a laser that emits a laser line normal to the x-y plane and two CCD cameras that detect the reflected laser line. The detected laser line is stored as intensity information for each pixel and, by

image processing and triangulation of this information, a 3D coordinate value is assigned for each measured point.

Motavalli and Valenzuela [13] used two fixed cameras with a 90° separation while a table on which the object is placed is rotated as shown in Fig 1.9.

Fig 1.9 Fixed camera/rotary table machine vision system [13]

Saint-Marc, Jezouin and Meioni [14] have designed a physical calibration gauge consisting of horizontal and vertical plates that, when fixed on the linear table, ensures that its axis of translation is perpendicular to the vertical plate or that, when fixed on the rotary table, ensures that its axis rotation lies into the vertical plate. Using two fixed cameras, a stereo image from the laser line is used to reconstruct the part surface Fig 1.10.

Fig 1.10 Stereo laser line reconstruction [14]

In a paper by Chang and Chang [15], a laser scanner is used as the object measurement system. When a laser line projects onto the object surface, the reflected laser image is acquired through two CCD devices. The hardware scheme of the laser scanner is shown in Fig 1.11. The laser scanner used can perform two types of scans, a plane scan and a rotated scan.

Fig 1.11 Linear and rotary laser scanner [15]

Campion, Dye, King and Ormson [16] are targeting multiple objects using a laser scanner. The basis of the system is a turn-table which supports four target mounts. As the turn-table rotates, the target mounts move radially towards or away from the center of the turn-table depending on the turn-table rotation as shown in Fig 1.12 and 1.13. This enables the focused laser spot to be scanned across a selected target by rotation of the turn-table. After each full rotation of the turn-table, the target mount has moved by about one spot diameter, then a different track across that target is addressed.

Fig 1.12 Top view, turn-table laser scanner [16]

Fig 1.13 Side view, turn-table laser scanner [16]

Pito [17] used laser stripes that rotate around the object collecting several samples to create the 3D image of the object. His algorithm uses the common strategy of keeping track of the “unseen” portions of the viewing volume to identify what needs to be scanned. An area of the viewing volume is “unseen” if it hasn’t been scanned due to omission or occlusion. First and second scans are shown in Fig 1.14 and 1.15.

Fig 1.14 Rotary scanner, showing voids [17]

Fig 1.15 Rotary scanner showing (a) first view (b) overlap [17]

Finally, Lee, Woo and Suk [18] classified the “Current Scan Mechanisms” devices based on the configuration of the machine. Although there are several configurations that have been explored, only three are dominantly used in industry. The three configurations are as follows:

Three Cartesian axis: as shown in Fig 1.16, the design follows the same structure as for CMMs. The inherent stability of the design allows for higher accuracy when compared to other configurations. By using a rotary table, an extra degree of freedom can be added.

Fig 1.16 Three Cartesian axis scanner [18]

Rotary platform on which the object is mounted with a stationary sensor: the sensor is traveling by rotating the platform and by lowering the platform in steps as shown in Fig 1.17.

Fig 1.17 Rotary scanner [18]

Robot arm with a sensor attached at the end of the arm: this type has a greater flexibility as it scans most shapes by moving the robot arm as shown in Fig 1.18.

Fig 1.18 Robotic arm scanner [18]

1.2 Proposed New Scanner

The current scanners have some disadvantages:

1. No single machine can scan six sides of a part, moreover the choice of a machine depends on the characteristics of the part to be scanned and its application.
2. A skilled operator is required to move the laser head around the part while maintaining fixed spacing between the sensor and the scanned object's surface.
3. An operator needs to plan a collision free path for the scanning device.
4. The configuration doesn't allow for bottom scanning.

As the automated system is one of the main problems in reverse engineering, a configuration that will reduce or minimize the number of occlusions must be selected.

It is for these reasons that a new scanner configuration, with the flexibility to move the scanner head in prismatic, cylindrical and spherical motions, was designed. The new design is shown in Fig 1.19.

Fig 1.19 Proposed scanner for orthogonal, cylindrical and spherical motions

The advantages of the proposed new scanner are:

1. Accessibility:
 - The Table can rotate up to 360° in the X-Y plane.
 - The CCD & laser can scan 360° in the Y-Z plane.
 - The glass table allows for bottom scan.
2. Automatic scanning: all points are within the depth of focus. Therefore, path planning is not required.
3. Non-contact laser allows scanning without surface damage of materials that are difficult to digitize on a CMM machine, e.g., clay, fabric, wood, foam, plaster, glass and flexible plastics.
4. Scanner can be built from "Off-the-shelf " parts.

Chapter 2

Construction of the Proposed Scanner

In order to study the merits of the proposed scanner, an experimental prototype was constructed at the Photonics Lab at Ryerson University. A 1200 x 2400 mm [48 x 96 in] optical table was used as the base of the prototype. Two stands built from melamine covered medium density fibreboard 305 mm [12 in] high and 127 mm [5 in] wide were constructed and mounted on to the optical table.

On the left hand side of the first stand (see Fig 2.1) a 330 mm [13 in] diameter plexiglass table and corresponding ball bearing are supported. The periphery of the rotary table has been precision CNC machined to fit an XL toothed belt. Driving the tooth belt is an Applied Motion from bipolar stepper motor. On the second stand, a pair of pillow blocks support a 19 mm [3/4 in] shaft, also driven by a stepper motor, geared with an anti-backlash gear with a 4.3:1 ratio to

provide more torque. Finally, at the end of the rotor arm, two linear guide rods and a lead screw support the prismatic axis of the line laser and CCD sled. A list of the mechanical components can be seen in Table 2.1.

Fig 2.1 AutoCAD drawing of the prototype

Table 2.1 Mechanical Components of Scanner

Name	Description
Motor 1	Applied motor stepper Pin # 5023-469
Motor 2	Applied motor stepper Pin # 5023-469
Motor 3	Applied motor stepper Pin # 5023-469
Rotary Table	Plexiglass with 208 teeth, 330 mm diameter
Drive Sprocket	XL tooth, delrin 13 teeth
Drive Belt	XL
Table Bearing	Ball Bearing, 330 mm diameter
Arm Shaft Bearing	Pillow Blocks, 19 mm
Drive Gear	tooth, 32 D.P.
Driven Gear	tooth, 32 D.P. anti-backlash
Lead Screw	Threaded rod 1/4 -20 UNC
Guide Rods	01 Drill rod, 6.35 mm diameter
Sled Nut	1/4 - 20 UNC, delrin

A photo of the prototype can be seen in Fig 2.2. The rotary table is surrounded with a toothed belt that allows motor #1 to rotate the platform. Motor #2 controls the movement of the sensor (laser diode and CCD camera) along the x-axis, through the means of the lead screw. The whole sensor can be swung around the plexiglass platform by motor #3. Therefore the following motions can be accomplished:

- Scan 360° around the object by rotating the table in the X – Y plane
- Scan the object 360° in the Y – Z plane by moving motor 3, carrying the sensor, in direction 3
- Scan the object along the X-axis by moving the sensor along the linear guides through the control of motor 2

Fig 2.2 A photo of the constructed scanner

With these movements, an object placed on the scanning table can be scanned with spherical, prismatic and cylindrical motions. As a result, a complete scan, including the bottom of the object through the clear table, can be performed.

2.1 Control of the Scanner

Motion is generated through three applied motion bi-polar stepper motors rated at 120 oz-in of torque. In Fig 1.19, along direction #1, the stepper motor provides a resolution of 0.6° , a resolution of $50\text{ }\mu\text{m}$ [0.00125 in] along direction #2 and a 2.3° resolution along direction #3. The stepper motors are all driven with Gecko G201 motor drivers. The drivers are interfaced with a PC, as shown in Fig 2.3, and are controlled through LabVIEW 5.1 with National Instrument's motion control board PCI-7324 and motion module UMI-7764. Fig 2.4 shows the picture of the motor control box.

Fig 2.3 PC interfacing with the scanner components

Fig 2.4 Motor control box

A LabVIEW virtual program has been written to control the scanner motion. Fig 2.5 shows the front panel and Fig 2.6 shows the LabVIEW block diagram to drive one motor. Implementing the program for all three motors, all motions of the scanner can be controlled. The motors calibrations are as follows:

- Motor 1, 360° represented by 590 steps
- Motor 2, 25.4 mm [1 in] represented by 800 steps
- Motor 3, 90° represented by 38 steps

Fig 2.5 Motor drive LabVIEW program (front panel)

Fig 2.6 Motor drive LabVIEW program (block diagram)

2.2 Scanning Sensor

The scanning sensor is comprised of a B&W CCD camera and a line projecting laser diode. Depth of the target object surface is calculated through triangulation. The camera and laser are placed 203 mm [8 in] apart, with the camera fixed perpendicular to the direction of travel of motor #2. The laser is pointed at 38° from the axis of the camera. The stand-off distance of the sensor has been selected at 254 mm [10 in].

A black and white (B&W) CCD camera is connected to a LabVIEW PCI-1407 image capture board. An f/8.5 mm lens is used to image the scene below the camera with a red filter to restrict the color of light reaching the CCD. The image size, 640x480 pixels is processed within LabVIEW using its IMAQ software. A LabVIEW virtual program has been written and can be seen in Fig 2.7 showing the front panel, and Fig 2.8 showing the block diagram. This program can capture one image and change it to a 640X480 array representing the illumination value of each pixel. The program also provides a 2X480 array, which represents the location of the strongest pixel in each row of 640 pixels.

Fig 2.7 Image capture LabVIEW program (front panel)

Fig 2.8 Image capture LabVIEW program (block diagram)

Adjusting the focus, illumination, and the alignment of the laser line are required before the scanner can be used to gather data. Adjusting focus is a three-step process. First, the LabView image capture program must be run in continuous mode so that a real time image is visible on the computer screen. The second step is to adjust the aperture to the maximum setting and the focus the CCD camera to get the best visual picture. As shown in Fig 2.9, a sheet of graph paper has been placed under the camera to aid in this adjustment process.

Fig 2.9 Focus adjustment using graph paper

Finally, the camera aperture is reduced to a setting where only the laser line is visible, as demonstrated in Fig 2.10. During this process, ambient room lights must be switched off to eliminate any reflection.

Fig 2.10 Adjustment of laser line illumination and alignment

To enhance the reading when more than one pixel having the same grey level, the following equation has been used [19]:

$$u = \sum u_i c_i / \sum c_i \quad (1)$$

Where i is from 1 to 640, c is the pixel position and u is the strength of the pixel (reading from the 640X480 array).

The black area has an illumination factor of 95 to 110 while the detected laser line starts from 120 to 255. To reduce noise and extraneous reflections pixels with grey level less than 115 have been thresholded.

2.3 Conversion of Pixel Values to XYZ Coordinates

In order to scan an object, the sensor (CCD and laser diode) sled is traversed in the X-direction with predetermined steps. At each step, a frame is taken by the CCD camera in which every 90 pixels represent 1 inch on the Y-axis, at the preset stand-off of the camera. Due to the scanner's mechanical construction, the origin ($y = 0$) is shifted by 188 pixels towards the Y-axis.

From the above, Y can be calculated as follows:

$$Y = [(\text{the reading in pixels in Y-direction}) - 188] / 90 \quad (2)$$

Triangulation is then used to calculate the height "B" of the target's surface by measuring the laser line shift registered on the CCD camera. Fig 2.13 is a photo of the actual sensor while the height is shown in Fig 2.14.

Fig 2.13 Camera/laser sled

Initially, the laser diode and CCD camera were built with a separation distance of 76 mm [3 in]. This resulted in a scanning resolution of 1.003 mm [0.0395 in] in the Z-direction and 0.302 mm [0.0119 in] in the X-direction. In order to increase the resolution and accuracy of the scanner the separation distance was increased to 203 mm [8 in]. The final resolution of the system is 0.381 mm [0.015 in] in Z-direction and 0.304 mm [0.012 in] in X-direction.

Although the resolution went slightly up from 0.302 mm [0.0119 in] to 0.304 mm [0.012 in] in X-direction, it is dramatically improved from 1.003 mm [0.0395 in] to 0.381 mm [0.015 in] in Z-direction.

Therefore, the X, Y and Z locations of each surface data point can be calculated as follows:

$$X = (\# \text{ of cycles of motor } \# 2/800) - 3 (\text{which is the shift in X to the centre}) - C \quad (8)$$

$$Y = [(\text{the reading in pixels in Y-direction}) - 188] / 90 \quad (9)$$

$$Z = 10 * (\text{shift in X-direction in pixels}) / (666 + \text{shift in X-direction in pixels}) \quad (10)$$

2.4 Transformation for the Rotary Table and Arm

Transformation to the original frame [20] is needed when the rotary table or the sensor arm is moved. Due to the rotation of the rotary table, the reference frame is rotated around the Z-axis by angle = θ

The transformation matrix is:

$$T_z = \begin{bmatrix} 1 & 0 & 0 \\ 0 & \cos \theta & -\sin \theta \\ 0 & \sin \theta & \cos \theta \end{bmatrix} \quad (11)$$

Due to the rotation of the arm carrying the sensor, the reference frame will rotate around the X-axis by angle = α . The transformation matrix will be

$$T_x = \begin{bmatrix} \cos \alpha & -\sin \alpha & 0 \\ \sin \alpha & \cos \alpha & 0 \\ 0 & 0 & 1 \end{bmatrix} \quad (12)$$

The final transformation matrix is the multiplication of the above two matrixes (11) & (12):

$$T = T_x * T_z = \begin{bmatrix} \cos \theta & -\sin \theta \cos \alpha & \sin \theta \sin \alpha \\ \sin \theta & \cos \theta \cos \alpha & -\cos \theta \sin \alpha \\ 0 & \sin \alpha & \cos \alpha \end{bmatrix} \quad (13)$$

Accordingly, if X' , Y' , Z' are calculated with respect to the camera frame and the X , Y , Z are calculated with respect to the machine reference frame, the relations will be as follows:

$$X = X' \cos \theta - Y' \sin \theta \cos \alpha + Z' \sin \theta \sin \alpha \quad (14)$$

$$Y = X' \sin \theta + Y' \cos \theta \cos \alpha - Z' \cos \theta \sin \alpha \quad (15)$$

$$Z = Y' \sin \alpha + Z' \cos \alpha \quad (16)$$

Where, the angle of rotation around the X and Z -axis can be calculated as follows:

$$\alpha^\circ = \text{X-axis rotation} = (\# \text{ of cycles in the loop command rotate motor 3}) * 90/38 \quad (17)$$

$$\theta^\circ = \text{Z-axis rotation} = (\# \text{ of cycles in the loop command rotate motor 1}) * 360/590 \quad (18)$$

2.5 Call Interface Node (CIN)

Using the CIN node is an alternative method for calling source code written in C language. The creation of the CIN node is done in three steps. The first step is generating a (.c) file by the CIN node from the LabVIEW program. Fig 2.15 and Fig 2.16 shows the (.c) files for cylindrical and spherical scans respectively. Adding the code needed where “/* ENTER YOUR CODE HERE */” sentence is shown. The second step is demonstrated in Appendix A, where a (.lsb) file is generated using Microsoft Visual C++ 6.0. The last step is downloading the (.lsb) file into the CIN node.

It must be noticed that the array format for the typedef struct in Fig 2.15 and Fig 2.16 are as follows:

```
typedef struct {  
    int32 dimSizes[2];  
    float64 arg1[1];  
} TD1;  
typedef TD1 **TD1Hdl;  
  
TD1Hdl var1;
```

According to the CIN user manual, accessing the element at row I and column J from an array with R rows and C columns, the following formulas are applied:

$$R = (* var1) \rightarrow \text{dimSizes} [0] \quad (19)$$

$$C = (* var1) \rightarrow \text{dimSizes} [1] \quad (20)$$

$$\text{Value} = (* var1) \rightarrow \text{arg1}[I * C + J] \quad (21)$$

```

/*
 * CIN source file
 */

#include "extcode.h"

/*
 * typedefs
 */

typedef struct {
    int32 dimSizes[2];
    float64 arg1[1];
} TD1;
typedef TD1 **TD1Hdl;

CIN MgErr CINRun(TD1Hdl var1, int32 *var2, float64 *Motor_1, float64
*Motor_3, float64 *Step);

CIN MgErr CINRun(TD1Hdl var1, int32 *var2, float64 *Motor_1, float64
*Motor_3, float64 *Step) {

    /* ENTER YOUR CODE HERE */

    return noErr;
}

```

Fig 2.15 (.c) file generated by CIN node for cylindrical scan

```

/*
 * CIN source file
 */

#include "extcode.h"

/*
 * typedefs
 */

typedef struct {
    int32 dim Sizes[2];
    float64 arg1[1];
} TD1;
typedef TD1 **TD1Hdl;

CIN MgErr CINRun(TD1Hdl var1, int32 *var2, float64
*Motor_3, float64 *Step);

CIN MgErr CINRun(TD1Hdl var1, int32 *var2, float64
*Motor_3, float64 *Step) {

    /* ENTER YOUR CODE HERE */

    return noErr;
}

```

Fig 2.16 (.c) file generated by CIN node for spherical scan

2.6 DXF Format

This format has been used for the purpose of displaying the result in a professional visual format. For simplicity, Autocad DXF format R 12 has been used [21]. Writing DXF format is performed in three steps. The first step is the beginning of the file, as shown in Fig 2.17, which is generated by the C code shown in Fig 2.18. The second step is displaying a point as shown in Fig 2.19, generated by the C code shown in Fig 2.20. The C code generating the points can be repeated as many times as the number of points to be displayed. The last step is the end of the file which is shown in Fig 2.21, generated by the C code shown in Fig 2.22.

```
0
SECTION
2
HEADER
0
ENDSEC
0
SECTION
2
TABLES
0
ENDSEC
0
SECTION
2
BLOCKS
0
ENDSEC
```

Fig 2.17 Beginning of DXF file format

```

if (first point)
{
    fprintf(OUTDXF,
        "0\nSECTION\n2\nHEADER\n0\nENDSEC");
    fprintf(OUTDXF,
        "\n0\nSECTION\n2\nTABLES\n0\nENDSEC");
    fprintf(OUTDXF,
        "\n0\nSECTION\n2\nBLOCKS\n0\nENDSEC");
}

```

Fig 2.18 C code to generate the beginning of DXF file

```

0
SECTION
2
ENTITIES
0
POINT
8
0
10
X coordinate
20
Y coordinate
30
Z coordinate
0
ENDSEC

```

Fig 2.19 Format for displaying a point by DXF file

```

fprintf(OUTDXF, "\n0\nSECTION\n2\nENTITIES\n0\nPOINT\n8\n0");
fprintf(OUTDXF, "\n10\n%f\n20\n%f\n30\n%f", X, Y, Z);
fprintf(OUTDXF, "\n0\nENDSEC");

```

Fig 2.20 C code to generate a point by the DXF file

```

0
EOF

```

Fig 2.21 End of DXF file format

```

If (last point)
{
    fprintf(OUTDAT, "\n0\nEOF\n");
}

```

Fig 2.22 C code to generate the end of DXF file

2.7 Master Algorithm

The overall master algorithm that controls the system motion and the CCD frame grabber generates two output files. The first contains the points collected from the object surface in text (TXT) format while the second file have the same points stored in AutoCAD (DXF) format.

A flow chart of the master algorithm for cylindrical scans is shown in Fig 2.23. The cylindrical scan program requires three inputs. The first input is the new angle of the rotary table by motor 1; the second is the rotation angle for the sensor arm in the Z – Y plane; and the third is the step length for the sensor in X-direction. The sensor must start from the home position.

A flow chart of the master algorithm for spherical scans is shown in Fig 2.24. The spherical scan program also requires two inputs. The first input is the rotation angle for the sensor arm in Z – Y plane while the second is the step for the rotary table angle. The sensor must start from the home position where $X = 0$ while the rotary table will rotate 360° in steps.

Fig 2.23 Cylindrical scan algorithm flowchart

Fig 2.24 Spherical scan algorithm flowchart

Chapter 3

Preliminary Results of the Prototype

A number of different gage blocks and angle blocks were used to test and calibrate the system. Cylindrical and spherical scans have been performed on a 1-2-3-gage block and a cast object as shown in Fig 3.1. The sensor was traversed at a set increment then an image was taken at each of those increments. The process was repeated until the complete surface was scanned. LabVIEW software was used to control the motion of the sensor, the capture and processing of the image set. To complete a scanning pass, approximately 5 minutes are required on a PII-400 MHz PC. The resulting scan data, typically with 5000 to 7000 points, can be saved in AutoCAD DXF format and brought into

AutoCAD for display purposes. The scan lines are 2.5 mm [0.1 in] apart with points spaced 0.3 mm [0.0125 in] apart within each line.

Fig 3.1 Cast object with quadric and free form surfaces

3.1 Laser Alignment and Separation

To achieve the alignment, a triangle consisting of the camera, the laser diode and the table, must be at a perfect right angle as previously demonstrated in Fig 2.14. To achieve the alignment, the zero reference ($z = x = 0$) is adjusted at 294 pixels from the image edge.

It is important for the laser diode to be fixed and positioned as vibrations from the sensor movement can affect its accuracy. To this end, the mounting bolts for the laser housing were tightened and worked with (white-out) to ensure that there was no movement from its original alignment.

After some preliminary scans, the accuracy of the scanner was improved from 1.003 mm [0.04 in] to 0.381 mm [0.015 in] by changing the camera laser distance from 76 mm [3 in] to 203 mm [8 in].

3.2 Motor Controls

All three motors play an important role in the image accuracy. When motor 2 is moving, the X coordinate is detected (see equation 8). When the transformation matrix is applied (see equation 13), the two angles θ and α are calculated from motors 1 and 3 respectively (see equation 17, 18).

The motor control is based on an open loop where there is no feedback. This is because the National Instrument's motion control board PCI-7324 allows for only four input/output, where three out of the four are configured as motor outputs.

Two calibrations have to be performed as a result of the open loop operation. The first is the home position, which must be done before each scan. The second is the step calibration, which is made every time the scanner is started.

Motors 1 and 3 move to the final destination in one step while motor 2 moves to the final destination in several steps. Accordingly, the motor calibrations are performed taking into consideration the inertia factor of the sled.

3.3 CIN Node

There is a specific method required to compile the CIN node. Starting with the LabVIEW program, a (.c) file is generated. Using the (.c) file as a base for a C++ compiler, specific commands shown in Appendix A must be followed to generate the complete (.lsb) file. Then, the (.lsb) file is downloaded into the CIN node in the LabVIEW program.

Familiarity with the data type is very critical, as interfacing between the labVIEW program and CIN node has a special format for data exchange. It must also be observed that using a CIN node customizes the LabVIEW program to fit a specific requirement.

It has also been observed that every time a change is made to the CIN program, the previous CIN node must be deleted and the computer must be booted before the new CIN program is installed. Otherwise, error messages will appear while running the program, as well as wrong results will be obtained.

As the CIN node is a type of Dynamic Link Library, all the errors will occur as run time errors where the CIN node has stopped the LabVIEW program.

3.4 Image Processing

In this section, scanning results are displayed on four-window figures using AutoCAD R 14. The top view on each figure is demonstrated on the upper left window, the isometric view on the lower left, the front view on the upper right, and the side view on the lower right.

In Fig 3.2, noise points are visible around the object, especially on the front and side view. This was caused by a reflective highlight on the polished steel 1-2-3 gage block. Rescanning of the block and subsequent scans were done after the gage blocks were spray painted with a light gray primer.

Fig 3.2 Reflections caused by shiny gage block

Figs 3.3, 3.4, and 3.5 show scans for the 1-2-3-gage block and the cast object (see Fig 3.1). Again, noise in the scans is seen under the block. A rough surface finish on the plexiglass table was thought to be the source of this erroneous data. A dulling gray (krylon dull-coat) was applied to the plexiglass in order to reduce laser reflections off the table surface. Although this reduced the amount of noise in each scan, the amount of noise was still not acceptable.

A new clearer plexiglass was machined with special care to protect the surface from scratches and abrasions. This helped to reduce the reflection from the background as only 5,000 to 7,000 points were collected from the object surface, instead of 20,000 to 25,000, where the difference is due to the plexiglass reflection. Storage memory was also reduced from 3M bytes to $\frac{1}{2}$ M bytes.

Fig 3.3 Gage scan with background reflection from the rotary table

Fig 3.4 Gage scan and reflection caused by the rotary table

Fig 3.5 Object scan and reflection caused by the rotary table

Figs 3.6 and 3.7 show reflections from ambient room light. Shutting off the room lights while the scan is performed, gives better results as shown in section 3.5.

Fig 3.6 1-2-3 –gage block scan with room light reflection

Fig 3.7 Cast object scan with room light reflection

The most sensitive adjustment in this test is the combination between the focus and aperture of the CCD when a bottom scan of the part is performed through the plexiglass. Two sets of reflections, off the bottom and top surfaces of the plexiglass, usually result. Adjusting the focus and aperture can reduce the intensity value of these reflections. Figs 3.8 and 3.9 are examples of unsuccessful bottom scans. Aperture settings and the focus were experimented with in order to get acceptable bottom scans.

Fig 3.8 Bottom scan with very low illumination

Fig 3.9 Bottom scan with high illumination

Finally, the rotary table is attached to the scanner body from one side. As a result, the weight of the scanned object can affect the table parallelism. Visibly, as shown in Fig 3.10, there is a slight skew to the resulting scan. One simple corrective action taken was to support the rotary table from both sides. However, this would limit the scanner arm from rotating the full 360° motion. Moreover, the side scan before applying the transformation matrix, is demonstrated on Fig 3.11.

Fig 3.10 Skew scan due to unsupported rotary table

Fig 3.11 Side scan, before applying the transformation matrix

3.5 Full Scan Results

A complete scan was achieved after the setup and adjustment had been performed. A 1-2-3-gage block and the cast object were scanned using the cylindrical and spherical methods.

3.5.1 Cylindrical Scan

A Cylindrical scan is performed by traversing the sensor on a set increment, done by motor 2 in direction 2, and taking an image at each increment. This was repeated until the complete surface was scanned. Fig 3.12 shows the machine adjusted for the top scan (arm at 90° from the home position). Figs 3.13 and 3.14 show the top scan of a 1-2-3-gage block in different poses whereas Figs 3.15, 3.16 and 3.17 show the arm in the same position as for the top scan while the rotary table is at 90° , 180° and 270° respectively. Fig 3.18 shows 4 different isometric views of the top scan for the cast object.

Fig 3.12 The machine positioned for a top scan

Fig 3.13 Top scan of a 1-2-3-gage block, where $Z = 1$

Fig 3.14 Top scan of a 1-2-3-gage block, where $Z = 3$

Fig 3.15 Top scan while the rotary table at 90°

Fig 3.16 Top scan while the rotary table at 180°

Fig 3.17 Top scan while the rotary table at 270°

Fig 3.18 Top scan for cast object with quadric and free form surfaces

Fig 3.19 shows the machine adjusted for a side scan, arm positioned 60° from the home position, while the rotary table is moved clockwise. The side scan of the 1-2-3-gage block is demonstrated by two different isometric views on the left hand side; while the right side shows the front view on the upper window and the side view on lower window. The casting object is demonstrated with three different isometric views. The top left corner always shows the top view.

Fig 3.19 The machine adjusted for the side scan

Figs 3.20, 3.21, 3.22 and 3.23 show the side scan for a 1-2-3-gage block where the rotary table is positioned at 0°, 90°, 180°, and 270° respectively. Figs 3.24, 3.25, 3.26 and 3.27 show the side scan (arm positioned 60° from the home position) for the cast object while the rotary table is positioned at 0°, 90°, 180°, and 270° respectively.

Fig 3.20 Side scan for 1-2-3-gage block while the rotary table at 0°

Fig 3.21 Side scan for 1-2-3-gage block while the rotary table at 90°

Fig 3.22 Side scan for 1-2-3-gage block while the rotary table at 180°

Fig 3.23 Side scan for 1-2-3-gage block while the rotary table at 270°

Fig 3.24 Side scan for cast object while the rotary table at 0°

Fig 3.25 Side scan for cast object while the rotary table at 90°

Fig 3.26 Side scan for cast object while the rotary table at 180°

Fig 3.27 Side scan for cast object while the rotary table at 270°

Fig 3.28 shows the machine adjusted for a bottom scan (arm at 270° from the home position). The bottom scan for the 1-2-3-gage block and the cast object is displayed on two different isometric views on the left side windows, while the right side has the front view on the upper window and the side view on lower window. Figs 3.29 and 3.30 show the bottom scan for a 1-2-3-gage block and the cast object respectively.

Fig 3.28 The machine adjusted for the bottom scan

Fig 3.29 Bottom scan for 1-2-3-gage block

Fig 3.30 Bottom scan for cast object

Based on the above results, it is obvious that the cylindrical scan can be used on both the top and the bottom. The percentage of noise points relative to the total points is close to zero on top scan, less than 0.5% on side scan and less than 2% on bottom scan. From the resulting point count, the noise points represent 0% when they were calculated for the experiment result shown in Fig 3.13, and 2% when they were calculated for the experiment result shown in Fig 3.29. As it was very difficult to count the points on the casting object, counting was done on the 1-2-3-gage block only.

3.5.2 Spherical Scan

A spherical scan is performed by traversing the rotary table a set distance by motor 1, and taking an image. This was repeated until the complete surface was scanned. Figs 3.31, 3.32 and 3.33 show spherical scans performed in a 1-2-3-gage block, while the arm is at 30° , 60° and 90° respectively from the home position. Figs 3.34 and 3.35 show the spherical scan performed on the cast object, while the arm is at 60° and 90° respectively from the home position.

Fig 3.31 Spherical scan for 1-2-3-gage block while the arm at 30°

Fig 3.32 Spherical scan for 1-2-3-gage block while the arm at 60°

Fig 3.33 Spherical scan for 1-2-3-gage block while the arm at 90°

Fig 3.34 Spherical scan for cast object while the arm at 60°

Fig 3.35 Spherical scan for cast object while the arm at 90°

When the spherical scan was performed, a top blank circle appeared with a radius of $r = 0.8 B$ (see equation 3, and Fig 2.14) where B is the height of the scanned object. As the sensor's arm angle was reduced from 90°, as shown in Fig 3.33, to 30°, as shown in Fig 3.31, the blank circle shrunk.

3.6 Calibration

Based on the above scans, the percentage of noise points relative to the total points is close to zero on top scan, less than 0.5% on side scan and less than 2% on bottom scan. This noise is detectable when using image segmentation software [1] as, in most cases it represents isolated points. Accordingly, top, side and bottom scans of the object have been successfully performed.

From the actual calibrations, the results are acceptable as the error in Y-axis is ± 0.289 mm [0.011 in], which represent $\pm$ one pixel in Y-direction and match the practical limitation. The error in Z-axis also matches the practical limitation, which is 0.381 mm [0.015 in] and represents $\pm$ one pixel. Due to the open loop in the motor controls, the error in X-direction does not match the practical limitation. It is ± 1.005 mm [0.04 in] (expected to be ± 0.354 mm [0.01325 in] which represents $\pm$ one pixel equal to ± 0.304 mm [0.012 in] added to the motor resolution which is ± 0.050 mm [0.00125 in]).

3.7 Comparison with Other Machines

Appendix F Technology Comparison Chart, shows tables for different types of scanners with their expected accuracies. By comparing the prototype to the line laser scanners, it is easy to realize that the prototype has respectable results.

The above-mentioned table shows the results of the sensors as well as the complete line laser scanners. The accuracy of the sensors are far more better than those of any complete line laser scanner, as the distance for measuring accuracy is less than one inch. The prototype has respectable resolution and accuracy compared to the complete line laser scanners, which have been introduced by Cyberware as shown in the same table.

Comparison of the bottom scan results is not applicable as the idea is new.

Finally, the advantage of accessibility and occlusion reduction makes the prototype unique in this area. The accuracy also can be improved to match the state-of-the art machines.

Chapter 4

Conclusions and Future Work

The proposed laser scanning system permits scanning of all sides of the object. The goals of the system are to allow for automated scanning for a wide variety of objects while minimizing the number of occlusions in the resulting scan data. From the results, the system provides scans on all sides of the object without the problems associated with reconstructing the surface from multiple scans, re-registration of the datum, or the time consuming problems of manual scanning.

Due to the top blank circle resulting in the spherical scan, the cylindrical scan is considered to have better surface information. However, the spherical scan can still be employed, as it is much faster with less adjustment. The spherical scan could also be used in combination with cylindrical scanning for special purposes.

It is recommended for resolution improvement to have a better resolution CCD camera, as well as DC motors with feed back control. Less reflection can be achieved by using an antireflection coating on both sides of the plexiglass rotary table. By using a bigger size arm and rotary table as well as a better resolution sensor, the restrictions on the scanned object dimensions 127x127x76 mm [5x5x3 in] can be improved.

Future work would include a surface fitting algorithm applied to the scanned data as well as a stereovision implementation. Automatic focus lenses can be employed with pre-set path to improve resolution as the high accuracy available in the market is attributed to close image capturing, which requires path planning. Additional constraints on the path planning algorithm must ensure that all surfaces being scanned are within the depth of range of the laser scanner and that the path has been optimized to reduce scan overlap.

References

- [1] V.H. Chan, C. Bradley, G.W. Vickers, A multi-sensor approach to automating co-ordinate measuring machine-based reverse engineering, *Computers in Industry*, 44, (2001)
- [2] Y. Takeuchi, H. Shimizu, I. Mukai, Automatic measurement of 3-dimensional coordinate measuring machine by means of CAD and image data, *Annals of the CIRP*, 39, (1990)
- [3] O. Ozeki, T. Nakano, and S. Yamamoto, Real-time range measurement device for three-dimensional object recognition, *IEEE Transactions on Pattern Analysis and Machine Intelligence*, PAMI-8, 4, (1986)
- [4] R.A. Jarvis, A Perspective on range finding techniques for computer vision, *IEEE Transactions on Pattern Analysis and Machine Intelligence*, PAMI-5, 2, (1983)
- [5] D.D. Harrison and M.P. Weir, High-speed triangulation-based 3-D imaging with orthonormal data projections and error detection, *IEEE Transactions on Pattern Analysis and Machine Intelligence*, 12, 4, (1990)
- [6] H. -Y. Feng, Y. Liu, F. Xi, Analysis of digitizing errors of a laser scanning system, *Precision Engineering, Journal of the International Societies for Precision Engineering and Nanotechnology*, 25, (2001)
- [7] F. Xi, Y. Liu and H.-Y. Feng, Error compensation for three-dimensional line laser scanning data, *Int. J Adv Manuf Technol*, 18 (2001)
- [8] J. Maver and R. Bajcsy, Occlusions as a guide for planning the next view, *IEE Transactions of Pattern Analysis and Machine Intelligence*, 15, 5,(1993)
- [9] J.J. Aguilar, F. Torres, M.A. Lope, Stereo vision for 3D measurement: accuracy analysis, calibration and industrial applications, *Measurement*, 18, 4, (1996)

- [10] M.W. Cho, M.K. Kim and K. Kim, Flexible inspection system based on a vision guided coordinate measuring machine, *Int. J. Prod. Res.*, 33, 5, (1995)
- [11] S. Son, S. Kim, K.H. Lee, Path planning of multi-patched freeform surfaces for laser scanning, *Int. J. Adv Manuf Technol*, 22, (2003)
- [12] K.H. Lee, H. Park and S. Son, A Framework for laser scan planning of freeform surfaces, *Int J Adv Manuf Technol*, 17, (2001)
- [13] S Motavalli and J. Valenzuela, A system for reverse engineering of prismatic parts using orthographic images, *Int. J. Computer Integrated Manufacturing*, 11, 2, (1998)
- [14] P. Saint-Marc, J. L. Jezouin, and G. Medioni, A Versatile PC-based range finding system, *IEEE Transactions on Robotics and Automation*, 7, 2, (1991)
- [15] D.-Y. Chang and Y.-M. Chang, A freeform surface modelling system based on laser scan data for reverse engineering, *Int J Adv Manuf Technol* 20, (2002)
- [16] R.P. Campion, D.R. Dye, P.J. King and R.G. Ormson, A multiple target system using laser ablation for multi-layer deposition, *Vacuum*, 47, 1, (1996)
- [17] R. Pito, A Solution to the next best view problem for automated surface acquisition, *IEEE Transactions On Pattern Analysis And Machine Intelligence*, 21, 10, (1999)
- [18] K. H. Lee, H. Woo and T. Suk, Data reduction methods for reverse engineering, *The Int. J. Adv. Manuf Technol*, 17, (2001)
- [19] M.A.G. Izquierdo, M.T. Sanchez, A. Ibanez and L.G. Ullate, Sub-pixel measurement of 3D surfaces by laser scanning, *Sensors and Actuators*, 76 (1999)
- [20] L. Sciavicco and B. Siciliano, *Modelling and control of robot manipulators*, Springer, ISBN 1852332212 (1999)
- [21] O. Ransen, *AutoCAD programming in C\C++*, Wiley, ISBN 0471963364 (1997)

Appendix A

Build the CIN Node (From CIN user manual, of the LabVIEW help menu)

To build CINs using Visual C++ Integrated Development Environment, complete the following steps:

- 1- From the LabVIEW program, generate (.c) file.
- 2- Create a new Dll project. Select File >> New >> select Win32 Dynamic-Link Library. Project name should be assigned at this point.
- 3- Add CIN objects and libraries to the project. Select Project >> Add To Project >> File and select cin.obj, labview.lib, lvsb.lib, lvsbmain.def from the Cintools subdirectory and (.c) file generated from step 1.
- 4- Add Cintools to the include path. Select Project >> Settings and change Setting For to All Configurations. Select the C/C++ tab and set the category to Preprocessor. Add the path to the Cintools directory in the Additional include directories field.
- 5- Set alignment to 1 byte. Select Project >> Settings and change Settings For to All Configurations. Select the C/C++ tab and set the category to Code Generation. Select the Struct member alignment tab and select 1 byte.
- 6- Choose a Run-time Library. Select Project >> Setting and change Setting For to All Configurations. Select the C/C++ tab and set the category to Code Generation. Select Multithreaded DLL in the Use Run-time Library control.
- 7- Make a custom build command to run lvsbutil. Select Project >> Settings and change Setting For to All Configurations. Select the Custom Build tab and change the Build command field as follows:
" <the path to cintools> \ lvsbutil" \$(TargetName) -d " \$ (WkspDir) \ \$ (OutDir)"
Change Output file field to
\$(OutDir) \$ (TargetName).lsb
- 8- Compile and Build All to generate (.lsb) file.
- 9- Load (.lsb) file to the CIN node in the LabVIEW program.

Appendix B

C Program for the CIN Node (Cylindrical)

```
/*
 * CIN source file
 */

#include "extcode.h"
#include <stdio.h>
#include <math.h>

/*
 * typedefs
 */

typedef struct {
    int32 dimSizes[2];
    int32 arg1[1];
} TD1;
typedef TD1 **TD1Hdl;

CIN MgErr CINRun(TD1Hdl var1, int32 *var2, float64 *Motor_1, float64 *Motor_3,
float64 *Step);

CIN MgErr CINRun(TD1Hdl var1, int32 *var2, float64 *Motor_1, float64 *Motor_3,
float64 *Step) {

    /* ENTER YOUR CODE HERE */

    FILE* OUTDXF;
    FILE* OUTDAT;
    int32 i ;
    double x, y, z, X , Y , z1 , Z , angz , angx ;
```

```

double pi = 3.1415926536 ;

OUTDXF = fopen("C:Test.dxf","a+");
OUTDAT = fopen("C:Test.dat","a+");

angz = (float) *Motor_1 * pi/180.0 ; /* z angle */
angx = (float) (*Motor_3 - 90) * pi /180.0 ; /* x angle */

if (*var2 == 0)
{
    fprintf(OUTDXF, "0\nSECTION\n2\nHEADER\n0\nENDSEC");
    fprintf(OUTDXF, "\n0\nSECTION\n2\nTABLES\n0\nENDSEC");
    fprintf(OUTDXF, "\n0\nSECTION\n2\nBLOCKS\n0\nENDSEC");
}

for (i=0 ; i < 480 ; i++)
{
    if ((*var1)->arg1[(2)*(i)+1] != 0)
    {
        x=(float) (*var2) * (*Step) - 2.5;
        y=(float)(188 - i)/90 ;
        z1= (float) 304 - (*var1)->arg1[(2)*(i)+1];
        x = (float) x - (z1+4.0)/88.0 ;
        z = (float)10*z1/(z1+252);
        X = x * cos (angz) - y * cos (angx) * sin (angz) + z * sin (angx) * sin (angz) ;
        Y = x * sin (angz) + y * cos (angx) * cos (angz) - z * sin (angx) * cos (angz) ;
        Z = y * sin (angx) + z * cos (angx) ;

        if ((Z>0)&&(Z<=3.5)&&(X>= -2.5)&&(X<=2.5)&&(Y>= -2.5)&&(Y<=2.5))
        {
            fprintf(OUTDAT, "%f, %f, %f\n",X,Y,Z);
            fprintf(OUTDXF, "\n0\nSECTION\n2\nENTITIES\n0\nPOINT\n8\n0");
            fprintf(OUTDXF, "\n10\n%f\n20\n%f\n30\n%f",X,Y,Z);
            fprintf(OUTDXF, "\n0\nENDSEC");
        }
    }
}

if (*var2 == ((int)((float) 5/(*Step)) - 1))
{
    fprintf(OUTDAT, "EOF\n");
}

```

```

        fprintf(OUTDXF, "\n0\nEOF\n");
    }

```

```

fclose(OUTDXF);
fclose(OUTDAT);

```

```

return noErr;
}

```

Appendix C

Full LabVIEW Program for Cylindrical Scan

Cylindrical_Final.vi

LL Snap in IMAQ Vision display acquires an image and displays it in an image window using the Low Level acquisition functions. A snap is appropriate for low-speed or single-capture applications where ease of programming is essential.

IMAQ Init is called to initialize the board. IMAQ Configure List configures the buffer list which contains one buffer. An image is created using IMAQ Create and wired to IMAQ Configure Buffer which configures the buffer in the buffer list. IMAQ Start starts the acquisition. IMAQ Get Buffer returns the image after it is acquired. This image is displayed in an image window using IMAQ WindDraw. IMAQ Close.vi is called to shut down the acquisition.

This example is designed to acquire 8-bit images. The IMAQ PCI-1424 can acquire from cameras with greater bit depths. If you are using a camera which outputs 10-, 12-, 14- or 16-bit images, change the image type input of IMAQ Create to 16 bits. If you are using a camera which outputs color images, change the image type input of IMAQ Create to RGB Chunky.

NOTE: You must have IMAQ Vision installed to run this example.

Connector Pane

Front Panel

For a description
select Show VI Info
from the Windows menu.

Interface Name

img0

Region of interest

0	0
0	0
0	32000
0	32000

NOTE: You must have IMAQ Vision
installed to run this example.

Array

300
200

Data is not available, as the
program is not running

size(s)

0

--

Numeric

0

Motor#1

0.00

Motor#3

0.00

Step

0.00

x+y

0.00

Array 2

200
0

Data is not
available, as
the program is
not running

Block Diagram

 

Appendix D

C Program for the CIN Node (Spherical)

```
/*
 * CIN source file
 */

#include "extcode.h"
#include<stdio.h>
#include<math.h>

/*
 * typedefs
 */

typedef struct {
    int32 dimSizes[2];
    float64 arg1[1];
} TD1;
typedef TD1 **TD1Hdl;

CIN MgErr CINRun(TD1Hdl var1, int32 *var2, float64 *Motor_3, float64
*Step);

CIN MgErr CINRun(TD1Hdl var1, int32 *var2, float64 *Motor_3, float64
*Step) {

    /* ENTER YOUR CODE HERE */

    FILE* OUTDXF;
    FILE* OUTDAT;
    int32 i ;
    double x, y, z, X , Y , z1 , Z , angz , angx ;
    double pi = 3.1415926536 ;

    OUTDXF = fopen("C:Test.dxf","a+");
    OUTDAT = fopen("C:Test.dat","a+");

    angz = (float) (*Step) * (*var2) * pi / 180 ; /* z angle */
}
```



```

angx = (float) (*Motor_3 - 90) * pi / 180 ; /* x angle */

if (*var2 == 0)
{
    fprintf(OUTDXF, "0\nSECTION\n2\nHEADER\n0\nENDSEC");
    fprintf(OUTDXF, "\n0\nSECTION\n2\nTABLES\n0\nENDSEC");
    fprintf(OUTDXF, "\n0\nSECTION\n2\nBLOCKS\n0\nENDSEC");
}

for (i=0 ; i < 480 ; i++)
{
    //x =(float) (*var2) * (*Step) - 2.5;
    y =(float)(188 - i)/91 ;
    z1= (float) 304 - (*var1)->arg1[(2)*(i)+1];
    z = (float)10*z1/(z1+666);
    //x = (float) 10.00 * z / 8.00;
    x = 0.00 ;
    X = x * cos (angz) - y * cos (angx) * sin (-angz) + z * sin
(angx) * sin (angz) ;
    Y = x * sin (-angz) + y * cos (angx) * cos (angz) - z * sin (-
angx) * cos (angz) ;
    Z = y * sin (-angx) + z * cos (angx) ;

    if ((Z> 0.02)&&(Z<=3.0))
    {
        fprintf(OUTDAT, "%f, %f , %f \n",X,Y,Z);
        fprintf(OUTDXF, "\n0\nSECTION\n2\nENTITIES\n0\nPOINT\n8\n0");
        fprintf(OUTDXF, "\n10\n%f\n20\n%f\n30\n%f",X,Y,Z);
        fprintf(OUTDXF, "\n0\nENDSEC");
    }
}

if (*var2 >= ((int)((float) 360/(*Step)) - 1))
{
    fprintf(OUTDAT, "EOF\n");
    fprintf(OUTDXF, "\n0\nEOF\n");
}

fclose (OUTDXF);
fclose (OUTDAT);

return noErr;
}

```

Appendix E

Full LabVIEW Program for Spherical Scan

Spherical_Final.vi

LL Snap in IMAQ Vision display acquires an image and displays it in an image window using the Low Level acquisition functions. A snap is appropriate for low-speed or single-capture applications where ease of programming is essential.

IMAQ Init is called to initialize the board. IMAQ Configure List configures the buffer list which contains one buffer. An image is created using IMAQ Create and wired to IMAQ Configure Buffer which configures the buffer in the buffer list. IMAQ Start starts the acquisition. IMAQ Get Buffer returns the image after it is acquired. This image is displayed in an image window using IMAQ WindDraw. IMAQ Close.vi is called to shut down the acquisition.

This example is designed to acquire 8-bit images. The IMAQ PCI-1424 can acquire from cameras with greater bit depths. If you are using a camera which outputs 10-, 12-, 14- or 16-bit images, change the image type input of IMAQ Create to 16 bits. If you are using a camera which outputs color images, change the image type input of IMAQ Create to RGB Chunky.

NOTE: You must have IMAQ Vision installed to run this example.

Connector Pane

Front Panel

For a description
select Show VI Info
from the Windows menu.

Interface Name

img0

Region of interest

0	0
0	0
0	32000
0	32000

NOTE: You must have IMAQ Vision
installed to run this example.

Array

300
200

Data is not available, as the
program is not running

size(s)

0

Numeric

0

Array 2

200
0

Data is not
available, as
the program is
not running

Motor #3

0.00

Step

0.00

x+y

0.00

Numeric 2

0

Appendix F

Technology Comparison Chart

Downloaded from "http://home.att.net/~castleisland/scn_08.htm"

This is by no means a complete listing. There are many more instruments available in each technology category and from each vendor, and from many additional vendors, as well. Entries were selected to illustrate the general specifications of a technology, or to show the upper and lower bounds of a vendor's offerings. Use the commercial directory pages to explore categories of interest in depth.

Technology	Representative Vendor	Model	Volume	Accuracy	Color ?	Speed	Price Range
Mechanical Touch Probe							
Mechanical Arm	Faro Technologies	FaroArm - Advantage	4 feet	0.001 inches	Not Applic	Manual operations	\$19,900
Mechanical Arm	Faro Technologies	FaroArm - Platinum	12 feet	0.005 inches	Not Applic	Manual operations	-
Mechanical Arm	Immersion	MicroScribe G2	50 inches (1.27 m)	0.015 inches (0.38 mm)	Not Applic	Manual operations	\$3,495
Mechanical Arm	Immersion	MicroScribe G2LX	66 inches (1.67 m)	0.012 inches (0.30 mm)	Not Applic	Manual operations	\$5,495
Guided Probe	Roland DGA Corp.	PIX-4	6 x 4 x 2.4 inches	Minimum scan pitch of 0.002 inches	Not Applic	No info provided	\$1,995
Guided Probe	Roland DGA Corp.	PIX-30	12 x 8 x 2.4 inches	Minimum scan pitch of 0.002 inches	Not Applic	No info provided	\$3,495
Strengths	accuracy; low-cost instruments avail.; measures deep slots, pockets; not affected by color or transparency						
Weaknesses	manual operation; slow for complex surfaces; can distort soft objects						

Technology	Representative Vendor	Model	Volume	Accuracy	Color ?	Speed	Price Range
<u>Laser Line Scanner</u>							
Laser Meas. Head	Laser Design, Inc.	RPS-120	Not Applic., Measuring head; 3.5 inches stand-off distance	+/- 0.00025 in acc.; 0.001 in resolution	No	14,400 points / sec	-
Laser Meas. Head	Laser Design, Inc.	RPS-450	Not Applic. Measuring head; 7.83 inches stand-off distance	+/- 0.001 in acc.; 0.0037 in resolution	No	14,400 points / sec	-
Laser Meas. Head	KREON Technologies	KLS 51	Not Applic. Measuring head; Mtd to Arm or Mach tool / Triangulation angle 30 deg	resolution 42 microns	No	Up to 20 000 points/sec	-
Laser Meas. Head	KREON Technologies	KLS 171	Not Applic. Measuring head; Mtd to Arm or Mach tool / Triangulation angle 30 deg	resolution 120 microns	No	Up to 20 000 points/sec	-
Laser Scanner / Aux. Video	Cyberware	Model 15 Desktop Scanner	9.84 (X) x 5.91 (min Y) x 2.95 (Z) inches	Resol: X,Y: 0.012; Z: 0.002 to 0.008 (inches)	Yes	14,580 points / sec.	-
Laser Scanner / Aux. Video	Cyberware	Model Shop 3D Scanner 3030RGB/MS	39 (X) x 10.25 (min Y) x 11.82 (Z) inches	Resol: X: 0.01 to 0.04; Y: 0.028; Z: 0.004 min (inches)	Yes	4,580 points / sec.	-
Laser Scanner / Aux. Video	Cyberware	Whole Body 4 color 3D Scanner	Cylindrical FOV; Depth: 47 inches; Height: 79 inches	Resol: X: 0.19; Y: 0.08; Z: 0.019 min (inches)	Yes	60,000 points / sec.	-

Laser Scanner / Aux. Video	Minolta	Vivid 700	2.8 x 2.8 to 43.3 x 43.3 inches	200 X 200 X 256 points (x,y,z resol.)	Yes	160,000 points in 0.6 sec.	-
Laser Scanner / Aux. Video	Minolta	Vivid 910	111 x 84 x 40 mm (tele lens) to 1200 x 903 x 400 mm (wide lens)	+/- 0.008 mm acc.; +/- 0.008 mm prec.	Yes	307,000 pixels in 2.5 secs using Fine mode and tele. lens	-
Laser Scanner	Nextec	Table Top Hawk	Linear travel: 240 mm (9.4 inch); Scanning range: +/- 5 mm (+/-0.2 inch)	Resolution: 1 micron; Total measuring accuracy: 10 microns (1 sigma)	No	40 points per sec	-
Laser Scanner / Aux. Video	Riegl	LMS-Z210i	4 to 400 m range x 80 deg vert x 360 deg Rot.	5 mm resol.; +/- 15 mm accuracy (averaged) or +/- 25 mm single shot	Yes, Optional	12,000 points / sec; up to 15 deg / sec for horiz scan	-
Laser Scanner	Riegl	LMS-Z420i	2 to 800 m range x 80 deg vert x 360 deg Rot.	5 mm resol.; +/- 5 mm accuracy (averaged) or +/- 10 mm single shot	No	12,000 points / sec; up to 15 deg / sec for horiz scan	-
Laser Scanner	Roland DGA Corp.	LPX-250	10 in (dia.) x 16 in (ht.)	0.008 inches resol.	No	No info provided	\$9,995
Laser Scanner	ShapeGrabber	LM600 System	600 x 160 x 165 mm	Accuracies to 0.03 mm	No	15,000 to 100,000+ points/sec	-
Strengths	non-contacting; fast digitizing of substantial volumes; good accuracy and resolution; color capability available						
Weaknesses	possible limitations for colored or transparent surfaces; laser cautions apply						

Technology	Representative Vendor	Model	Volume	Accuracy	Color ?	Speed	Price Range
Other Types of Laser Systems							
Laser time of flight	Callidus Precision Systems GmbH	CT900	1.6 m W x 1.4 m H	50 microns resolution on axis of turntable; 70 microns at far end of meas range	No	4000 points / sec	-
Laser time of flight	Callidus Precision Systems GmbH	CT180	350 mm W x 375 mm H	25 at near end to 70 microns at far end of meas range	No	4000 points / sec	-
Laser tracking	Automated Precision, Inc.	Smart Trak 6D	40 m range	Static acc: 0.001 inches at 16 ft	No	Up to 2000 meas / second tracking a moving target	-
Strengths	non-contacting; fast digitizing of substantial volumes; good accuracy and resolution; trackers are very precise; large stand-off distance for large objects						
Weaknesses	possible limitations for colored or transparent surfaces; trackers are fairly slow and may req targets; laser cautions apply						

Technology	Representative Vendor	Model	Volume	Accuracy	Color ?	Speed	Price Range
<u>Tracking Systems, Other Than Laser Types</u>							
LED-based tracking	Boulder Innovation Group, Inc.	3D Creator	1 m Sphere	0.2 mm mean volumetric acc	No	300 points / second	-
LED-based tracking	Metronor	Solo Measurement System	10.5 m range	0.02 to 0.1 mm meas. uncertainty	No	Manual operations	-
Magnetic tracking	Northern Digital	Aurora	500 mm cubic volume	1.3 mm rms; 2.1 mm to 95% confid. level	No	22 to 45 points / sec depending on number of sensor coils	-
Strengths	can digitize hidden points in some cases; can be used to track motion of many points simult.; magnetic devices can meas outside line of sight; not affected by color, transparency or surface quality						
Weaknesses	requires targets or contacting probes; smaller volumes; lower accuracy; slow for complex surfaces						

Technology	Representative Vendor	Model	Volume	Accuracy	Color ?	Speed	Price Range
Structured-light and Broadband Source Systems							
Arm-mounted structured white light scanners	GOM mbH	ATOS II	5 x 4 x 4 to 67 x 53 x 53 inches	0.003 to 0.039 inches	No	1.3 million points in 7 secs.	-
Arm-mounted structured white light scanners	GOM mbH	ATOS III	6 x 6 x 4 to 80 x 80 x 80 inches	0.003 to 0.040 inches	No	4 million points in 8 secs.	-
Moire white light pattern	Inspeck	3D Mega Capturor II / Small Field	401 x 321 mm	X & Y: 0.3 mm; Z: 0.4 mm resol.	Yes	1.3 million points in 0.7 sec.	-
Moire white light pattern	Inspeck	3D Mega Capturor II / Large Field	1140 x 910 mm	X & Y: 0.9 mm; Z: 1.0 mm resol.	Yes	1.3 million points in 0.7 sec.	-
Moire white light pattern	Inspeck	Capturor II / Small Field	352 x 264 mm	X & Y: 0.6 mm; Z: 0.5 mm resol.	Yes	0.3 million points in 0.4 sec.	-
Moire white light pattern	Inspeck	Capturor II / Large Field	1195 x 896 mm	X & Y: 1.9 mm; Z: 1.0 mm resol.	Yes	0.3 million points in 0.4 sec.	-
Projected white light patterns	Genex Technologies, Inc.	Rainbow 3D Camera	Several customizable models offered with capacities from approximately a 3 inch cube to more than a 10 inch cube.	Accuracy to 25 microns (0.001 in)	Yes	Captures 768 X 576 pixel image (442,368 pixels) in < 1 sec.	-
Moire-coded white light triangulation	Steinbichler Optotechnik GmbH	Comet	Not specified; 2 million points per view	Accuracy to +/- 20 microns	No	Not specied	-
Strengths	very fast; color available; no laser safety precautions						
Weaknesses	somewhat lower accuracy						

Technology	Representative Vendor	Model	Volume	Accuracy	Color ?	Speed	Price Range
Internal Viewing Systems							
X-ray computed tomography	Aracor	Konoscope 160/200	200 mm dia	1024 X 1024 X 1024 pixel image; 0.4 mm resol; 0.04 mm acc.	Not Applic	No info provided	-
X-ray computed tomography	Aracor	ICT 2500	2500 mm dia	2048 X 2048 pixels; 3mm resol; 0.25 mm acc.	Not Applic	No info provided	-
X-ray computed tomography	Bio-Imaging Research (BIR)	Actis 600/450	600 mm dia; unlimited height	20 line pairs/ mm max spatial resol; density resol: 0.1%; can detect 150 micron flaw, 200 micron void and 10 micron crack.	Not Applic	14 min for scanning 600 mm dia object	-
X-ray computed tomography	HYTEC Sensors & Imaging Group, Inc.	Flash CT	16 inches in height; cross sectional area undefined	0.005 inches resolution	Not Applic	200 MB / minute of volume data	-
Destructive / Cross-sectional Scanning	CGI, Inc.	CSS-300	7 x 5 x 9 inches	+/- 0.0008 inches (0.0005 to 0.010 inches layer thickness)	No	2 to 3 hours	-
Destructive / Cross-sectional Scanning	CGI, Inc.	CSS-3000	14 x 16 x 15 inches	+/- 0.0008 inches (0.0005 to 0.010 inches layer thickness)	No	No info provided	-

Strengths	Internal viewing and sample preservation for CT						
Weaknesses	radiation safety; sample destruction for CSS; can be slow						
Technology	Representative Vendor	Model	Volume	Accuracy	Color ?	Speed	Price Range